

## United States Patent and Trademark Office

UNITED STATES DEPARTMENT OF COMMERCE United States Patent and Trademark Office Address: COMMISSIONER FOR PATENTS P.O. Box 1450 Alexandria, Virginia 22313-1450 www.uspto.gov

PPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO
10/632,028 07/31/2003		Andrew J. Ries	P9173.00	9663
27581 75	590 02/28/2006		EXAMINER	
MEDTRONIC, INC.			ALTER, ALYSSA M	
710 MEDTRONIC PARK MINNEAPOLIS, MN 55432-9924			ART UNIT	PAPER NUMBER
MINNEM ODIO, MIN OD 102 372			3762	

DATE MAILED: 02/28/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

	_	
- 1		-
£		
•		

Application No.	Applicant(s)		
10/632,028	RIES ET AL.		
Examiner	Art Unit		
Alyssa M. Alter	3762		

Advisory Action	10/632,028	RIES ET AL.				
Before the Filing of an Appeal Brief	Examiner	Art Unit				
	Alyssa M. Alter	3762				
The MAILING DATE of this communication appe	ars on the cover sheet with the c	orrespondence add	ress			
THE REPLY FILED 06 February 2006 FAILS TO PLACE THIS	APPLICATION IN CONDITION FO	OR ALLOWANCE.				
<ol> <li>The reply was filed after a final rejection, but prior to or on the same day as filing a Notice of Appeal. To avoid abandonment of this application, applicant must timely file one of the following replies: (1) an amendment, affidavit, or other evidence, which places the application in condition for allowance; (2) a Notice of Appeal (with appeal fee) in compliance with 37 CFR 41.31; or (3) a Request for Continued Examination (RCE) in compliance with 37 CFR 1.114. The reply must be filed within one of the following time periods:</li> <li>a) The period for reply expiresmonths from the mailing date of the final rejection.</li> </ol>						
b) The period for reply expires on: (1) the mailing date of this Advisory Action, or (2) the date set forth in the final rejection, whichever is later. In no event, however, will the statutory period for reply expire later than SIX MONTHS from the mailing date of the final rejection.  Examiner Note: If box 1 is checked, check either box (a) or (b). ONLY CHECK BOX (b) WHEN THE FIRST REPLY WAS FILED WITHIN TWO						
MONTHS OF THE FINAL REJECTION. See MPEP 706.07(f	).					
Extensions of time may be obtained under 37 CFR 1.136(a). The date on been filed is the date for purposes of determining the period of extension a CFR 1.17(a) is calculated from: (1) the expiration date of the shortened sta above, if checked. Any reply received by the Office later than three month earned patent term adjustment. See 37 CFR 1.704(b).  NOTICE OF APPEAL	and the corresponding amount of the fee. atutory period for reply originally set in the safter the mailing date of the final rejection	The appropriate extension final Office action; or (2) on, even if timely filed, may	n fee under 37 as set forth in (b) y reduce any			
2. The Notice of Appeal was filed on A brief in compliance with 37 CFR 41.37 must be filed within two months of the date of filing the Notice of Appeal (37 CFR 41.37(a)), or any extension thereof (37 CFR 41.37(e)), to avoid dismissal of the appeal. Since a Notice of Appeal has been filed, any reply must be filed within the time period set forth in 37 CFR 41.37(a).  AMENDMENTS						
<u> </u>	but prior to the date of filing a brie	f will not be entered b	necalise			
(a) They raise new issues that would require further consideration and/or search (see NOTE below);						
(b) ☐ They raise the issue of new matter (see NOTE belo (c) ☐ They are not deemed to place the application in be		educing or simplifying	the issues for			
appeal; and/or (d)☐ They present additional claims without canceling a		jected claims.				
NOTE: (See 37 CFR 1.116 and 41.33(a)).						
4. The amendments are not in compliance with 37 CFR 1.121. See attached Notice of Non-Compliant Amendment (PTOL-324).  5. Applicant's reply has overcome the following rejection(s):						
6. Newly proposed or amended claim(s) would be allowable if submitted in a separate, timely filed amendment canceling						
the non-allowable claim(s).  7. For purposes of appeal, the proposed amendment(s): a) will not be entered, or b) will be entered and an explanation of how the new or amended claims would be rejected is provided below or appended.						
The status of the claim(s) is (or will be) as follows:  Claim(s) allowed:						
Claim(s) objected to: Claim(s) rejected:						
Claim(s) withdrawn from consideration:						
AFFIDAVIT OR OTHER EVIDENCE						
<ol> <li>The affidavit or other evidence filed after a final action, b because applicant failed to provide a showing of good ar and was not earlier presented. See 37 CFR 1.116(e).</li> </ol>						
9. The affidavit or other evidence filed after the date of filing a Notice of Appeal, but prior to the date of filing a brief, will <u>not</u> be entered because the affidavit or other evidence failed to overcome <u>all</u> rejections under appeal and/or appellant fails to provide a showing a good and sufficient reasons why it is necessary and was not earlier presented. See 37 CFR 41.33(d)(1).						
10. ☐ The affidavit or other evidence is entered. An explanation REQUEST FOR RECONSIDERATION/OTHER	on of the status of the claims after e	entry is below or attac	hed.			
11. The request for reconsideration has been considered by See Continuation Sheet.	ut does NOT place the application i	n condition for allowa	nce because:			
12. Note the attached Information Disclosure Statement(s). (PTO/SB/08 or PTO-1449) Paper No(s).						
13. Other:						
		GEORGE R. EVAL PRIMARY EXAM	NISKO IMED			
		2/3/6	1EM			

Continuation of 11. does NOT place the application in condition for allowance because: the Applicant argues that Lim (US 5,769,671) does not discloses a deflection portion deflecting the connector clip from a first position to a second postion. However, in accordance with the previous Office Actions, the examiner explained that the first postion of the spring is prior to the springs insertion into the housing. Once the spring is inserted into the housing, the housing maintains the spring at a second postion different from the first postion. The spring is inherently slightly deflected in the housing at second position compared to before placement in the housing inorder to facilitate sufficent contact with the housing and the lead. Therefore, Lim does in fact teach a deflection portion for deflecting the connector clip from a first position to a second position.

Alyssa M. Alter